

# EU Quality Management System Certificate TW23/00000625

The management system of

## Microlife Corporation

The SGS logo consists of the letters 'SGS' in a bold, black, sans-serif font. A thin orange horizontal line is positioned below the letters, and a thin orange vertical line is positioned to the right of the letters, intersecting the horizontal line.

9F, No. 431, RuiGuang Road, NeiHu, Taipei, 114, Taiwan, R.O.C.  
SRN Number: TW-MF-000010688

has been assessed and certified as meeting the requirements of

### MDR EU Quality Management System certificate (Annex IX QMS)

For the following products

**The Scope of Registration appears on page 2 of this certificate**

This certificate is valid from 05 April 2024 until 24 August 2028 and remains valid subject to satisfactory surveillance audits.

Re certification audit due before 24 February 2028

Issue 3. Certified since 24 August 2023

Certified activities performed by additional sites are listed on subsequent pages.

A handwritten signature in blue ink, appearing to be 'VS', is written over a horizontal line.

Authorised by

Virginie Siloret

Global Medical Device

Certification Manager

SGS Belgium NV NB1639

SGS House Noorderlaan 87 2030 Antwerp Belgium

t +32 (0)3 545-48-48 - www.sgs.com

This document is an authentic electronic certificate for Client' business purposes use only. Printed version of the electronic certificate are permitted and will be considered as a copy. This document is issued by the Company subject to SGS General Conditions of certification services available on [Terms and Conditions](#) | SGS. Attention is drawn to the limitation of liability, indemnification and jurisdictional clauses contained therein. This document is copyright protected and any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful.



EU Quality Management System Certificate TW23/00000625,  
continued

# Microlife Corporation

## MDR EU Quality Management System certificate (Annex IX QMS)

Class IIa - MDA0203, MDS1010

Blood Pressure Long-term Ambulatory Recorder Series

Basic UDI-DI:  
4719003ABPBC

Digital Non-invasive Blood Pressure Monitor for Home Use Series

Basic UDI-DI:  
4719003HBPCF

Digital Infrared Thermometer Series

Basic UDI-DI:  
4719003IRSR

Digital Thermometer Series

Basic UDI-DI:  
4719003MTT9-

Digital Non-invasive Blood Pressure Monitors for Professional Use Series

Basic UDI-DI:  
4719003OBPDJ

Digital Peak Flow Meters Series

Basic UDI-DI:  
4719003PFSN

Conditions for & limitation to the validity of the certificate:

For placing on the market of Class III or class IIb implantable devices (except sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors and Annex VIII rule 12 devices) covered by this certificate, a Technical Documentation Assessment Certificate according to Annex IX section 4 and 5 is required.

For Class I devices, audit done by SGS Belgium N.V. is limited to the specific aspect described in Article 52 section 7 of MDR (EU) 2017/745 (sterility, reusability or measurement function).

List of examinations and tests performed, which may include reference to relevant CS and harmonised standards, as per Annex XII, Chapter II, section 10 is available "on request" per email to [NB1639@sgs.com](mailto:NB1639@sgs.com).

Limitation: N/A

Certification is based on following reports: - TW/TPE/613065 - CTC 1.11

Authorized representative Name and address (if relevant): Microlife UAB P. Lukšio g. 32 08222 Vilnius, Lithuania

Previous certificate number: N/A

Change in between this certificate and previous one: Scope rewording and adding Digital Non-invasive Blood Pressure Monitors for Professional Use Series and Digital Peak Flow Meters Series

This document is an authentic electronic certificate for Client' business purposes use only. Printed version of the electronic certificate are permitted and will be considered as a copy. This document is issued by the Company subject to SGS General Conditions of certification services available on [Terms and Conditions](#) | SGS. Attention is drawn to the limitation of liability, indemnification and jurisdictional clauses contained therein. This document is copyright protected and any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful.



**SGS**

EU Quality Management System Certificate TW23/00000625,  
continued

# Microlife Corporation

The SGS logo is positioned in the top right corner. It features the letters 'SGS' in a bold, black, sans-serif font. A thin orange horizontal line is positioned below the letters, and a thin blue vertical line is positioned to the right of the letters, intersecting the orange line.

## MDR EU Quality Management System certificate (Annex IX QMS)

Issue 3

### Sites

Microlife Corporation  
9F, No. 431, RuiGuang Road, NeiHu, Taipei, 114, Taiwan, R.O.C.

ONBO Electronic (Shenzhen) Co., Ltd.  
No.138, Huasheng Road, Langkou Community, Dalang Street, Longhua District, Shenzhen, China.

This document is an authentic electronic certificate for Client' business purposes use only. Printed version of the electronic certificate are permitted and will be considered as a copy. This document is issued by the Company subject to SGS General Conditions of certification services available on [Terms and Conditions](#) | SGS. Attention is drawn to the limitation of liability, indemnification and jurisdictional clauses contained therein. This document is copyright protected and any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful.

